



CERTIFICATE OF ACCREDITATION

The ANSI National Accreditation Board

Hereby attests that

International Wheel and Tire Inc.

**23255 Commerce Drive
Farmington Hills, MI 48335**

Fulfills the requirements of

ISO/IEC 17025:2017

In the fields of

CALIBRATION and DIMENSIONAL MEASUREMENT

This certificate is valid only when accompanied by a current scope of accreditation document.
The current scope of accreditation can be verified at www.anab.org.

A handwritten signature in black ink, appearing to be 'Jason Stine', is positioned above a horizontal line.

Jason Stine, Vice President

Expiry Date: 24 March 2027

Certificate Number: ACDM-3188



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017.
This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory
quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).

SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

International Wheel and Tire Inc.

23255 Commerce Drive
Farmington Hills, MI 48335
Guru Gopalan
248-298-0207

CALIBRATION

Valid to: **March 24, 2027**

Certificate Number: **ACDM-3188**

Mass and Mass Related

Parameter	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Weights for Master Wheels	0 g to 600 g	0.014 g	Comparison to ASTM E617 Class 2 weights using Electronic Weigh Scale 0.01 g Resolution

DIMENSIONAL MEASUREMENT


3 Dimensional

Parameter	Range	Expanded Uncertainty of Measurement (+/-)	Reference Standard, Method, and/or Equipment
Dimensional measurement (Set-up rings; Master Wheel) 1D, 2D, 3D	X Axis Up to 700 mm Y Axis Up to 1 000 mm Z Axis Up to 450 mm	0.03 mm	Measurement using Coordinate Measuring Machine as Reference Standard

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 ($k=2$), corresponding to a confidence level of approximately 95%.

Notes:

1. This scope is formatted as part of a single document including Certificate of Accreditation No. ACDM-3188.



Jason Stine, Vice President